




AQ-SPEC
Air Quality Sensor Performance Evaluation Center

AQ-SPEC: SENSOR PERFORMANCE EVALUATIONS

BRANDON FEENSTRA

DAVIS, CA - 06/05/19

PRIMARY QUALITY ASSURANCE ORGANIZATION TRAINING



OUTLINE

BACKGROUND

AQ-SPEC

- LAB
- FIELD
- FIELD PM SENSOR RESULTS




HISTORICAL AMBIENT AIR MONITORING


Clean Air Act (CAA)

- Approved Federal Reference (FRM) and Equivalent Methods (FEM)
- Regional Networks = Determine Regional Attainment of NAAQS

Recent legislation

- CA Assembly Bill AB617 (Passed)
- House Bill 1284 - Crowd Sourcing of Environmental Data Act of 2019 (Introduced)





AQ-SPEC

Air Quality Sensor Performance Evaluation Center

LOW-COST SENSORS

- Technical advancements
- Cost Reductions
 - Hardware
 - Connectivity
 - Cloud computing
- Rapidly gaining attention


POTENTIAL FOR AMBIENT AIR MONITORING

- Fence-line, community, hot-spot identification, mobile monitoring, personal exposure, science education



1800-1980

2011



2012

2013: EPA Next Generation Air Sensors Conference

2014



2015-2019: New Start-ups

2016-2019: Development of sensor networks



HOW CAN SENSORS FIT IN?

Public Health

- What are the concentrations?
- What is the spatial-temporal variability?

Inform Mitigation Efforts

- What are the sources and their relative contribution?
- How much is locally produced versus transported over long distances?

Inform and Empower the public


- Collect data applicable to personal exposure
- Modify habits to reduce exposure



AQ-SPEC

Air Quality Sensor Performance Evaluation Center

- Established in July 2014
 - Over \$600,000 initial investment
- Main Goals & Objectives
 - Provide guidance & clarity
 - Promote successful evolution and use of sensor technology
 - Minimize confusion
- Sensor Selection Criteria
 - Commercially available
 - Criteria pollutants & air toxics
 - Real- or near-real time, time resolution \leq 5-min
 - Sensitivity at ambient levels
 - Continuous operation for two months
 - Retrievable data
 - Low-cost...?





AQ-SPEC

Air Quality Sensor Performance Evaluation Center

FIELD PERFORMANCE EVALUATION

- Sensors tested in triplicates
- Two month deployment (various time intervals, random)
- South Coast AQMD Riverside-Rubidoux Air Monitoring Station
 - Fully instrumented
 - Inland site
 - 1 km from CA SR 60
 - Impacted by regional aerosol formation





AQ-SPEC

Air Quality Sensor Performance Evaluation Center

LABORATORY PERFORMANCE EVALUATION



- ✓ Outer chamber
- ✓ Made of stainless steel
- ✓ Shape: Rectangular
- ✓ Volume: 1.3 m³
- ✓ HVAC system
- ✓ Louvered ceiling surface
- ✓ Set of two fans



- ✓ Inner chamber
- ✓ Teflon-coated Stainless Steel
- ✓ Shape: Cylindrical
- ✓ Volume: 0.11 m³



AQ-SPEC

Air Quality Sensor Performance Evaluation Center

LABORATORY EVALUATION (HISTORY AND WHAT'S NEXT)

Received Chamber Summer 2015

- AQ-SPEC staff worked closely with the chamber system integrator to customize the system to the anticipated needs and requirements (12 months)
- AQ-SPEC staff developed Methods for Aerosol and Gas testing atmospheres in-house (6 months)

PAPAPOSTOLOU V, ZHANG H, FEENSTRA B, AND POLIDORI A. [DEVELOPMENT OF AN ENVIRONMENTAL CHAMBER FOR EVALUATING THE PERFORMANCE OF LOW-COST AIR QUALITY SENSORS UNDER CONTROLLED CONDITIONS](#). *ATMOSPHERIC ENVIRONMENT*, 171: 82-90, 2017

State-of-the-art system

Systematically evaluate performance of sensors

Produce stable and reproducible PM and Gas test atmospheres

Produce a wide range of known target/interferent pollutant concentrations,
temperature and relative humidity conditions

Sensor data communication options



AQ-SPEC

Air Quality Sensor Performance Evaluation Center

EVALUATION PARAMETERS:

- Intra-model variability
- Accuracy
- Precision
- Coefficient of Determination (R^2)
- Data Recovery
- Climate Susceptibility
- Interferent (monodisperse aerosols)

16 Lab Evaluations Completed

www.aqmd.gov/aq-spec/evaluations/laboratory

Challenges:

- Stability of PM_{10} atmospheres
 - Due to nature of test particles
- Sensor performance degradation experiments
- Temperature and RH cycling tests for long periods of time



AQ-SPEC

Air Quality Sensor Performance Evaluation Center

- Developing:
 - ASTM D22.05 Testing Protocol
 - Low-Cost Indoor Air Quality sensors for measuring CO₂ and PM_{2.5}
 - VOC Sensor Testing Protocol
 - Total VOC, speciated VOC
- New contract in place for a second chamber system (delivered end of 2019) to accommodate testing of 20+ sensors simultaneously, aging/vibration/wind effects/rapid climatic changes experiments:
 - Development of sensor performance standards
 - AQ-SPEC Sensor Library program
 - Testing protocol for sensors conducting mobile ambient air measurements



PM SENSORS – FIELD EVALUATION RESULTS

In Review: Feenstra, et al. 2019. Performance Evaluation of Twelve Low-cost PM_{2.5} Sensors at an Ambient Air Monitoring Site, *Atmospheric Environment*

PERFORMANCE EVALUATION PARAMETERS

- Intra-model variability
- Accuracy
- Measurement Error
- Impact of local conditions

DATA FILTERS TO IMPROVE INTER AND INTRA-MODEL COMPARISON

- PM_{2.5} > 50 µg/m³ removed
- If reference or any of 3 sensors missing a 1-hr value, data row removed from analysis



PERFORMANCE EVALUATION OF TWELVE LOW-COST PM_{2.5} SENSORS AT AN AMBIENT AIR MONITORING SITE

Manufacturer	Model	Pollutants Measured	Time Resolution	Cost
Aeroqual	AQY	PM _{2.5} , O ₃ , NO ₂	1-min	\$3,000
Airboxlab	Foobot	PM _{2.5} , CO ₂ , VOC	5-min	\$200
Alphasense	OPC-N2	PM _{2.5}	< 1-min	\$450
HabitatMap	Air Beam 1	PM _{2.5}	1-min	\$200
Hanvon	N1	PM _{2.5} , HCHO	1-min	\$200
Kaiterra	LaserEgg	PM _{2.5}	< 1-min	\$200
PurpleAir	PA-II	PM _{2.5} , PM ₁₀ , PM _{1.0}	< 1-min	\$230
SainSmart	Pure Morning P3	PM _{2.5} , CO ₂ , HCHO	< 1-min	\$170
Shinyei	PM Evaluation Kit	PM _{2.5}	1-min	\$1,000
TSI	AirAssure	PM _{2.5}	5-min	\$1,000
Uhoo	uhoo	PM _{2.5} , O ₃ , NO ₂ , CO, CO ₂ , TVOC	1-min	\$300
IQAir	AirVisual Pro	PM _{2.5} , CO ₂	< 1-min	\$270



BIAS ERROR CALCULATIONS

$$\text{Mean Bias Error (MBE)} = \frac{1}{n} \sum_{i=1}^n (X_i - X_t)$$

$$\text{Mean Absolute Error (MAE)} = \frac{1}{n} \sum_{i=1}^n |X_i - X_t|$$

Where,

X_i is the 1-hr average measurement by the low-cost sensor

X_t is the 1-hr average measurement provided by the reference

n is the number of 1-hr time-matched data pairs



REGRESSION AND MEASUREMENT ERROR

AEROQUAL

- Intercept and Bias

HANVON & PURPLE AIR & SAINSMART

- Over-estimate

KAITERRA LASER EGG

- Importance of R²


UHOO

- Poor performance

Sensor	#	R ²	Slope		Intercept		Measurement Error (µg/m ³)	
			Slope	95% CI	Intercept	95% CI	MBE	MAE
Aeroqual AQY	1	0.78	0.99	0.02	-2.75	0.39	-2.9	4.5
	2	0.79	1.01	0.02	-3.08	0.38	-3.0	4.7
	3	0.79	0.94	0.02	-2.63	0.35	-3.4	4.6
Airboxlab Foobot	1	0.57	1.32	0.06	0.28	1.00	5.0	6.4
	2	0.54	1.08	0.05	1.35	0.86	2.6	4.7
	3	0.54	1.29	0.07	4.89	1.03	9.2	9.5
Alphasense OPC	1	0.67	0.78	0.04	2.08	0.67	-1.3	3.3
	2	0.38	0.57	0.05	1.18	0.90	-5.5	6.5
	3	0.40	0.67	0.06	1.03	1.01	-4.2	5.9
HabitatMap Air Beam 1	1	0.59	1.08	0.05	2.03	0.63	2.9	4.4
	2	0.57	1.47	0.07	0.46	0.90	5.7	6.5
	3	0.57	1.66	0.08	-0.62	1.01	6.8	7.5
Hanvon N1	1	0.56	2.13	0.10	0.91	1.71	17.4	18.1
	2	0.54	1.91	0.10	2.69	1.59	15.9	16.3
	3	0.58	1.73	0.08	2.39	1.34	13.1	13.5
Kaiterra LaserEgg	1	0.57	1.15	0.06	-0.08	0.95	2.0	4.7
	2	0.56	1.02	0.06	-0.40	0.85	-0.1	4.1
	3	0.58	1.01	0.06	-0.80	0.82	-0.7	4.0
PurpleAir PA-II	1	0.95	1.68	0.03	-3.06	0.51	5.0	7.0
	2	0.95	1.63	0.03	-2.84	0.49	4.7	6.7
	3	0.95	1.58	0.03	-2.08	0.48	4.8	6.7
SainSmart Pure Morning P3	1	0.76	1.52	0.05	-2.34	0.69	3.5	5.3
	2	0.77	1.61	0.05	-2.19	0.70	4.6	5.9
	3	0.74	1.31	0.05	0.06	0.62	3.5	5.0
Shinyei PM Evaluation Kit	1	0.75	1.18	0.04	-1.48	0.59	0.9	4.5
	2	0.73	1.13	0.04	-1.07	0.60	0.7	4.5
	3	0.75	1.03	0.03	-1.29	0.52	-0.9	4.2
TSI AirAssure	1	0.73	1.10	0.04	1.61	0.60	2.9	5.1
	2	0.74	1.08	0.03	3.66	0.57	4.7	6.0
	3	0.72	1.01	0.03	3.81	0.56	4.0	5.6
Uhoo	1	0.00	0.09	0.11	31.11	2.03	15.4	17.7
	2	-	-	-	-	-	-	-
	3	0.00	0.02	0.08	19.74	1.51	2.9	10.1
IQAir AirVisual Pro	1	0.69	1.15	0.04	-2.38	0.73	0.2	4.4
	2	0.69	1.16	0.04	-2.42	0.73	0.3	4.4
	3	0.72	1.31	0.04	-1.97	0.77	3.4	5.3


RH AND BIAS ERROR

- BAM is equipped with heater
- Sensors measure at ambient
- Some sensors correct for RH bias
- Typically, see increasing positive bias error as RH increases



PM CONCENTRATION AND BIAS ERROR

- SYSTEMATIC VS RANDOM ERROR





AQ-SPEC

Air Quality Sensor Performance Evaluation Center

THANK YOU - QUESTIONS?

AQ-SPEC Team

Jason Low
Andrea Polidori
Vasileios Papapostolou
Brandon Feenstra
Berj Der Boghossian
Wilton Mui
Ashley Collier-Oxandale
Michelle Kuang
Steve Boddeker


DATA RECOVERY

Sensor		Temp (°C)	RH (%)	BAM PM _{2.5} (µg/m ³)		Data Recovery (%)		
Manufacturer & Model	Evaluation Dates	Mean ± SD	Mean ± SD	Mean ± SD	Max	BAM	Sensor*	Analysis (N) **
Aeroqual AQY	12/22/17 - 03/27/18	14.9 ± 5.6	48.2 ± 27.1	13.8 ± 14.4	133	88	99	84 (1917)
Airboxlab Foobot	07/14/16 - 09/15/16	25.2 ± 5.7	53.1 ± 21.6	14.4 ± 6.4	38	96	95	86 (1295)
Alphasense OPC-N2	07/10/15 - 08/10/15	24.7 ± 4.9	58.8 ± 19.5	15.6 ± 6.6	45	99	99	98 (732)
HabitatMap Air Beam 1	03/17/17 - 05/12/17	18.1 ± 5.3	53.5 ± 23.2	11.1 ± 6.6	47	98	99	98 (1317)
Hanvon N1	05/20/16 - 07/27/16	23.5 ± 6.7	54.1 ± 22.0	15.2 ± 10.3	131	98	88	77 (1264)
Kaiterra LaserEgg	08/01/16 - 09/26/16	24.2 ± 5.8	54.6 ± 21.6	14.0 ± 6.1	38	96	92	71 (951)
PurpleAir PA-II	12/08/16 - 01/26/17	12.3 ± 4.0	67.9 ± 25.3	12.1 ± 11.3	73	97	99	96 (1124)
SainSmart Pure Morning P3	03/17/17 - 05/12/17	18.1 ± 5.3	53.5 ± 23.2	11.1 ± 6.6	47	99	93	78 (1047)
Shinyei PM Evaluation Kit	02/05/15 - 04/08/15	18.0 ± 6.1	48.1 ± 26.3	15.2 ± 12.3	79	99	99	97 (1435)
TSI AirAssure	12/18/15 - 02/15/16	13.5 ± 5.7	47.6 ± 27.3	13.2 ± 11.3	69	96	93	91 (1299)
Uhoo	08/07/17 - 10/06/17	24.2 ± 6.3	55.7 ± 21.5	17.1 ± 7.3	51	99	79***	92 (1333)
IQAir Air Visual Pro	08/02/17 - 10/05/17	24.5 ± 6.2	55.9 ± 21.0	17.2 ± 7.3	51	99	99	98 (1535)
Mean of Means ± SD		20.1 ± 4.6	54.3 ± 5.3	14.2 ± 2.0				

SUMMARY STATISTICS AND INTRA-MODEL VARIABILITY FOR SENSOR TRIPLICATES

	Sensor Mean \pm SD ($\mu\text{g}/\text{m}^3$)			Mean of Means
Sensor	1	2	3	Mean \pm SD ($\mu\text{g}/\text{m}^3$)
Aeroqual AQY	9.8 \pm 11.5	9.7 \pm 11.7	9.3 \pm 10.8	9.6 \pm 0.24
Airboxlab Foobot	19.7 \pm 10.3	17.3 \pm 8.6	24.0 \pm 10.3	20.3 \pm 2.75
Alphasense OPC-N2	14.3 \pm 6.2	10.1 \pm 6.1	11.4 \pm 7.0	11.9 \pm 1.74
HabitatMap Air Beam 1	14.1 \pm 9.3	17.0 \pm 12.8	18.0 \pm 14.5	16.4 \pm 1.64
Hanvon N1	32.0 \pm 21.7	30.5 \pm 19.7	27.6 \pm 17.3	30.0 \pm 1.80
Kaiterra LaserEgg	15.6 \pm 9.2	13.5 \pm 8.2	12.9 \pm 8.0	14.0 \pm 1.16
PurpleAir PA-II	16.9 \pm 19.1	16.5 \pm 18.6	16.7 \pm 18.0	16.3 \pm 0.13
SainSmart Pure Morning P3	14.6 \pm 12.2	15.7 \pm 12.8	14.7 \pm 10.6	15.0 \pm 0.51
Shinyei PM Evaluation Kit	14.8 \pm 13.1	14.6 \pm 12.7	13.0 \pm 11.5	14.1 \pm 0.80
TSI AirAssure	15.6 \pm 13.4	17.4 \pm 13.0	16.7 \pm 12.4	16.6 \pm 0.75
Uhoo	32.6 \pm 14.9	-	20.1 \pm 11.0	26.3 \pm 6.23
IQAir Air Visual Pro	17.5 \pm 10.2	17.6 \pm 10.2	20.7 \pm 11.4	18.6 \pm 1.51

RUBIDOUX SEASONAL CHEMICAL COMPOSITION OF PM_{2.5}



Seasonal average chemical composition of PM_{2.5} between 2002 and 2013 at Rubidoux monitoring station. Data adapted from Hasheminassab et al. (2014).

OPC METHODOLOGY

Detects by:

Outputs:

Possible issues:

Type 1: Optical Particle Counter

Sizing individual particles by how they scatter light; counts particles per size bin (e.g., 16 size channels); converts to a "mass" concentration based on assuming particles are spheres and have a certain density.

Manufacturer dependent.
Possibilities include:
-Count per size bin
-Estimated mass for PM_{1} , $PM_{2.5}$, PM_{10}

- Under-counting at high concentrations
- Bias due to assumptions (e.g., density)
- Measurement artifacts

Type 2: "Total scattering" type sensor

Particles as a group scatter light – this is converted to an estimated concentration (e.g. mass or number of particles per unit volume)

Usually a single numeric output: voltage, calculated concentration

- Upper and lower detection limit issues
- Potential bias due to big particles (e.g., $>10 \mu m$)