SEPA 2023 O_3 TAD and **Cross-Section Requirement Updates**

PQAO Training –

February 27-28, 2024

Amy Seeds Kai Crombie EPA Region 9

Introductions and Overview

Amy Seeds – EPA Region 9

Monitoring and Analysis Section (MAS) Kai Crombie – EPA Region 9 Quality Assurance Branch (QAB)



Introductions and Overview

Amy Seeds – EPA Region 9

Monitoring and Analysis Section (MAS) Kai Crombie – EPA Region 9 Quality Assurance Branch (QAB)

Part 1: Ozone (O₃) Technical Assistance Document (TAD) Updates Part 2:

Ozone Absorption Cross-Section Value Requirement Update



O₃ TAD General Organization

- Section 1 Introduction
- Section 2 EPA's O₃ Traceability Scheme
- Section 3 Types of O₃ Devices
- Section 4 Verification and Reverification Requirements for O_3 Transfer Standards
- Section 5 Verification and Reverification Procedure for O_3 Transfer Standards
- Section 6 Operational Considerations of O₃ Transfer Standards





Continued...

- Appendix A Equations and Example Calculations
- Appendix B Example Acceptance Testing Data Sheet
- Appendix C Utilizing the National Institute of Standards and Technology (NIST) 7 Essential Elements of Traceability
- Appendix D Rationale and Testing Methodology for O_3 Verification and Reverification Acceptance Criteria
- Appendix E Qualification Testing



Section 1 - Introduction

2013 Version (EPA – 454/B-13-004)	2023 Version (EPA-454/B-22-003)
Transfer standard (TS) nomenclature was ambiguous to application and distance from Level 1	Transfer standard (TS) nomenclature is clarified and based on application and distance from Level 1
Reverification frequency was based on distance from standard reference photometer (SRP)	Reverification is based on application
Relative Standard Deviation was used as a measure of TS verification AND reverification cycles stability	Standard Deviation is used for <u>verification</u> cycles stability. Change from previous regression slope and intercept are used for <u>reverification</u> stability.
6 cycles were required for verifications and cycles had to be on different days	3 stables cycles are required for verification and can be conducted on same day
Allowed generator-only devices to be used as O ₃ TS	Does NOT allow use of generator-only devices as O ₃ TS
Level 4 TS were allowed	Level 4 TS are strongly discouraged and are allowed only if additional requirements are met
Did not include best practices	Includes hands-on operational best practices for working with O_3 TS















U.S. Environmental Protection Agency

Section 2 – Standards Definition

Transfer Standard Levels

Defined by distance from SRP

- Level 1 = SRP
- Level 2 = compared to SRP
- Level 3 = compared to Level 2

Bench Standard

- Traditional "primary Traditional "field standard"

Field Standard

- Stationary Standards
 Moves from site to site
 - standard"



NIST SRP pictured above

https://www.nist.gov/programs-projects/nist-standardreference-photometer-ozone-measurement-traceability



Section 2 – EPA's Traceability Scheme Figure 1 – Traceability Pyramid Level 1 (SRP) Increased Uncertainty is Level 2 controlled by following (Verified to L1) established procedures Level 3 (Verified to L2) O₃ Monitor at Monitoring Site

Increasing Number of Transfer Standards



Section 2 – Verification Rules and Frequency



* Indicates verification annually** Indicates verification every 6 months

Verification timeframe begins on the day the verification is completed. Not the day the standard is put into service. (Section 2.6)



U.S. Environmental Protection Agency

Section 2 – <u>Reverification</u> Rules and Frequency



A TS may not be used after the reverification frequency timeperiod has been exceeded. However, it may not require a new verification (all 3 cycles). See Section 2.7



U.S. Environmental Protection Agency

Section 3 – Types of O₃ Transfer Standards

2013 TAD

2023 TAD

Allowed generatoronly devices to be used as O₃ Transfer Standard



Section 3 – Types of O₃ Transfer Standards

2013 TAD

2023 TAD

Allowed generatoronly devices to be used as O₃ Transfer Standard



Section 3 – Types of O₃ Transfer Standards

2013 TAD

2023 TAD

Allowed generatoronly devices to be used as O₃ Transfer Standard Does not allow generatoronly devices to be used as O_3 Transfer Standard



Section 4 – Verification and Reverification Requirements for O₃ Transfer Standards 2013 TAD 2023 TAD

Levels had different acceptance criteria

Level 4 Transfer Standards were allowed

6 cycles were required for verifications and had to be on different days



Section 4 – Verification and Reverification Requirements for O₃ Transfer Standards 2013 TAD 2023 TAD

Levels had different acceptance criteria

Level 4 Transfer Standards were allowed

6 cycles were required for verifications and had to be on different days



Section 4 – Verification and Reverification Requirements for O₃ Transfer Standards 2013 TAD 2023 TAD

Levels had different acceptance criteria

Level 4 Transfer Standards were allowed

6 cycles were required for verifications and had to be on different days Levels 2 and 3 have the same acceptance criteria

Level 4 TS are strongly discouraged and are only allowed if additional criteria are met

3 cycles are required for verification and can be performed on the same day

Section 4.2.2 – Acceptance Testing Requirements

- Acceptance testing is required as part of normal practice for verification/reverification
 - Must be completed...
 - after a new TS is received from the manufacturer
 - Prior to verification or reverification of a TS
 - When a TS is shipped
 - When a device requires repair
- Example form in Appendix B



Section 4 – Acceptance Criteria

- Same acceptance criteria for levels 2 and 3
 - Verification (Equations = Appendix A)
 - Each point difference must be < ±3.1% (Equation 1) or ±1.5 ppb for concentration points below 50 ppb (Equation 2).
 - All 3 regression slopes must be 1.00 ± 0.03 (Equation 4)
 - All 3 regression intercepts must be 0 ± 3 ppb (Equation 4)
 - SD of the 3 slopes must be < 0.0075 (Equation 8)
 - SD of the 3 intercepts must be < 1.00 ppb (Equation 9)

Transfer Standards for Calibration of Air Monitoring Analyzers for Ozone, Technical Assistance Document (January 2023)



Section 4 – Acceptance Criteria

- Reverification (Equations = Appendix A)
 - Use the data collected during the reverification cycle to calculate the following:
 - Percent difference (% Diff) in measured concentration, calculated at each tested concentration point > 50 ppb and must be within ±3.1% (Equation 1)
 - Absolute difference (*AbsDiff*) in measured concentration, calculated at each tested concentration point ≤ 50 ppb must be within ±1.5 ppb (Equation 2)
 - The calculated value of the regression slope (*m*) must fall within 1.00 ± 0.03 (i.e., within ±3% of 1.00) (Equation 4)
 - m must **ALSO** fall within ±0.015 of the calculated mean slope (\overline{m}) from the most recent successful verification test (Equation 6)
 - The calculated value of the regression intercept (b) must fall within 0.00 ± 0.03 ppb (Equation 4)
 - *b* must **ALSO** fall within ±1.5 ppb of calculated mean intercept (\overline{b}) from the most recent successful verification test (Equation 7)



O₃ TAD Implementation

November 15, 2023 – November 15, 2025

- TAD was published on November 15, 2023
- "implementation will occur over a phase in period of not more than 2 years from the time this document becomes final" ¹

¹ Transfer Standards for Calibration of Air Monitoring Analyzers for Ozone, Technical Assistance Document (January 2023)



U.S. Environmental Protection Agency

Part 2

Ozone Absorption Cross-Section Value Requirement Update



What is the Absorption Cross-Section for Surface Ozone Measurements?

 $(4)^2$

Absorption cross-section (absorption coefficient, α) used to determine atmospheric O₃ concentrations based on the amount of light absorbed at an ultraviolet (UV) wavelength of 253.65 nm

Transmittance
$$= \frac{I}{I_0} = e^{-\alpha c l}$$
 (1)²

$$[O_3]_{OUT} = \left(\frac{-1}{\alpha l} \ln \frac{I}{I_0}\right) \left(\frac{T}{273}\right) \left(\frac{760}{P}\right) \times \frac{10^6}{L}$$

Where:

 $[O_3]_{OUT} = O_3$ conc., ppml = optical path length, cm $P = sample pressure, torr<math>\alpha = absorption coefficient<math>T = sample temp, K$ $L = O_3 loss correction factorof <math>O_3$ @ STP, atm⁻¹ cm⁻¹factorfactor



² Appendix D to Part 50—Reference Measurement Principle and Calibration Procedure for the Measurement of Ozone in the Atmosphere (Chemiluminescence Method), 4.1 Principle



U.S. Environmental Protection Agency

Cross-Section Value Requirement Update

- Updated value represents a more accurate and precise value
- It improves the accuracy of UV O₃ analyzers and SRPs
- Consensus value is 1.23% lower and uncertainty 4.25x smaller
- On October 12, 2023, the EPA revised 40 CFR Part 50 App D -Reference Measurement Principle and Calibration Procedure for the Measurement of Ozone in the Atmosphere (Chemiluminescence Method)
- Federal Register Notice posted on <u>EPA AMTIC website</u>

 $308 atm^{-1}cm^{-1}$ "current" or "old" value Uncertainty of $4 atm^{-1}cm^{-1}$ (1.4%)

Reference Measurement Principle and Calibration Procedure for the Measurement of Ozone in the Atmosphere (Chemiluminescence Method), Federal Register Volume 88, Issue 196 (October 12, 2023)



Cross-Section Value Requirement Update

- Updated value represents a more accurate and precise value
- It improves the accuracy of UV O₃ analyzers and SRPs
- Consensus value is 1.23% lower and uncertainty 4.25x smaller
- On October 12, 2023, the EPA revised 40 CFR Part 50 App D -Reference Measurement Principle and Calibration Procedure for the Measurement of Ozone in the Atmosphere (Chemiluminescence Method)
- Federal Register Notice posted on <u>EPA AMTIC website</u>
- The update will be achieved through software/firmware modification and will NOT require any hardware changes

Reference Measurement Principle and Calibration Procedure for the Measurement of Ozone in the Atmosphere (Chemiluminescence Method), Federal Register Volume 88, Issue 196 (October 12, 2023)

 $308 atm^{-1}cm^{-1}$ "current" or "old" value Uncertainty of $4 atm^{-1}cm^{-1}$ (1.4%) $304.39 atm^{-1}cm^{-1}$ "updated" value Uncertainty of 0.94 $atm^{-1}cm^{-1}$ (0.31%)



Managing and Implementing the Update



 $304.39 atm^{-1}cm^{-1}$ "updated" value







 $304.39 atm^{-1}cm^{-1}$ "updated" value



Managing and Implementing the Update



 $304.39 atm^{-1}cm^{-1}$ "updated" value



Managing and Implementing the Update



 $304.39 atm^{-1}cm^{-1}$ "updated" value



Questions?

Amy Seeds – EPA Region 9 seeds.amy@epa.gov Kai Crombie – EPA Region 9 crombie.kai@epa.gov

Part 1: Ozone (O₃) Technical Assistance Document (TAD) Requirement Updates Part 2: Ozone Absorption Cross-Section Value Requirement Update

